

1. Technical Data

| General | |
|---|---|
| Dimensions | 11.3"W x 7.2"H x 2.4"D (288 mm x 182 mm x 60 mm) |
| Weight | 4.5 lb (2.0 g) |
| Finish | PC/ABS |
| Power Requirements | 100 – 240 VAC, 1.2 A max |
| Mains Frequency Range | 50 – 60 Hz |
| Power Consumption | 12W nominal, 30W (when charging battery) |
| Standards Conformance | IEC 60601-1:2005 (General Safety) AAMI ES60601-1:2005 (General Safety) CSA C22.2#60601-1:2008 (General Safety) IEC 60601-1-2:2007 (Class B) (EMC) IEC 60601-1-4:2000 (General Safety) IEC 60601-1-8:2006 (Alarms) IEC 60601-2-27:2011 (ECG Monitoring) AAMI SP10:2002/A1:2003 (Non-Invasive Blood Pressure) ISO 80601-2-30:2009 (Non-Invasive Blood Pressure) IEC 60601-2-34:2011 (Invasive Blood Pressure) ISO 80601-2-55:2011 (CO ₂ Respiratory Gas Monitoring) ISO 80601-2-56:2009 (Temperature) ISO 80601-2-61:2011 (SpO ₂) IEC 62366:2007 (Usability) IEC 62304:2006 (Software) |
| Patient Risk Current (IEC 60601-1) | Electro Veterinary Apparatus with Isolated Patient Connection. Meets the following limits: Enclosure Risk Current < 100 µA Patient-applied Risk Current < 10 µA Patient Isolation Risk Current < 50 µA Earth Risk Current < 500 µA |
| Type of Protection (Electrical) | Class I |
| Degree of Protection (Electrical) | Type CF, Defibrillation-proof |
| Degree of Protection (Water) | Ordinary Equipment (IPX1) |
| Disinfecting Method | Per the instructions in the Cleaning chapter |
| Degree of Safety (Flammable Anesthetic Mixture) | Not suitable for use in the presence of a Flammable Anesthetic Mixture |
| Mode of Operation | Continuous |
| | |
| | |
| | |

| Battery | |
|-----------------------------|--|
| Type | Lithium-Ion Rechargeable |
| Discharging Time | 4 hours (minimum) |
| Charging Time | 5 hours |
| Charging Method | Battery is charged while monitor is connected to AC main |
| Environmental | |
| Cooling | Convection (no fan) |
| Operating Temperature | 32 to 104 °F (0 to 40 °C) |
| Storage Temperature | -4 to 140 °F (-20 to 60 °C) |
| Operating Humidity | 15% to 90% non-condensing |
| Storage Humidity | 15% to 95% non-condensing |
| Operating Altitude | 0 to 15,000' (0 to 4572 m) |
| Storage Altitude | 0 to 40,000' (0 to 12,192 m) |
| Alarm Signal Sound Pressure | 45 to 80 dB(A) |
| Display | |
| Type | Active Matrix LCD |
| Size | 8.4 inches (diagonal) |
| Matrix | 800 x 600 pixels |
| Number of Waveform Channels | Up to 5 |
| Sweep Speed | 6.25, 12.5, 25 mm/s |
| Display Mode | Eraser Bar |
| ECG | |
| Accessories | 3-lead cable, 5-lead cable |
| Input Connector | 7-pin connector |
| Displayable Leads | 3-lead cable: I, II, III, AVL, AVR, AVF 5-lead cable: I, II, III, AVL, AVR, AVF, V |
| HR Resolution | 1 bpm (beats per minute) |
| Measurement Range | 15 to 300 bpm |
| Measurement Accuracy | ±2 bpm or ±1%, whichever is greater |
| Response Time | Per IEC 60601-2-27, change from 80 to 120 bpm: < 7 seconds Per IEC 60601-2-27, change from 80 to 40 bpm: < 11 seconds |
| Report Interval | 1 second |

| | |
|----------------------------------|---|
| HR Averaging Scheme | Average of the 10 most recent, valid R-R intervals, discarding the shortest and longest interval |
| Time to Alarm - Tachycardia | IEC 60601-2-27, ECG Complex B1: < 10 sec (5 sec typical) IEC 60601-2-27, ECG Complex B2: < 10 sec (9 sec typical) |
| Notch Filter Frequency | 50Hz, 60 Hz, Off |
| Filter Bandwidth | Monitor Mode: 0.67 Hz to 40 Hz (-3 dB) Diagnostic Mode: 0.05 to 40 Hz (-3 dB) |
| Dynamic Range AC | ±5 mV, per IEC 60601-2-27 |
| Dynamic Range DC | ±300 mV, per IEC 60601-2-27 |
| Defibrillation Protection | Complies with IEC 60601-2-27 |
| Pacer Pulse Detection | Lead II, I and V |
| Pacer Pulse Rejection | Rejects all pulses of amplitude ±2mV to ±700mV and duration 0.1 to 2 ms, per IEC 60601-2-27, Clause 201.12.1.101.13 |
| Tall T-Wave Rejection | Rejects T-Waves less than or equal to 120% of a 1mV QRS and a Q-T interval of 350 ms, per IEC 60601-2-27, Clause 201.12.1.101.17 |
| HR Response to Irregular Rhythm | IEC 60601-2-27, ECG Complex A1: HR is 80 bpm IEC 60601-2-27, ECG Complex A2: HR is 65 bpm IEC 60601-2-27, ECG Complex A3: HR is 120 bpm IEC 60601-2-27, ECG Complex A4: HR is 91 bpm |
| Active Noise Suppression | RL drive (< 5 µA) |
| Pulse Tone | Yes |
| Pulse Oximetry | |
| Method | |
| Input Connector | Absorption – Spectrophotometric (dual wavelength) (Functional oxygen saturation of arterial hemoglobin) |
| SpO ₂ / PR Resolution | 9-pin connector |
| Measurement Range | SpO ₂ : 1 O ₂ % PR: 1 bpm (beat per minute) |
| Measurement Accuracy | SpO ₂ : 20 to 100% PR: 30 to 240 bpm |
| Measurement Test Method | SpO ₂ : from 70 to 100%: ±2% (O ₂ %), < 70%: unspecified PR: ± 3 bpm |
| Report Interval | Comparison versus co-oximeter, per ISO 80601-2-61 |
| Pulse Tone | 1 second. Numeric values held < 30 seconds |
| | |
| | |

| Non-Invasive Blood Pressure | |
|---|--|
| Method | Oscillometric |
| Input Connector | Single Lumen Hose (Quick-Disconnect fitting) |
| Cuff | Small and Large Animal, |
| Derived Parameters | Systolic, Mean, Diastolic |
| Resolution | 1 mmHg |
| Measurement Range | Systolic: 30 to 250 mmHg Mean: 20 to 230 mmHg Diastolic: 10 to 210 mmHg |
| Measurement Accuracy | Complies with AAMI SP10 |
| Transducer Accuracy | ± 3 mmHg |
| Pulse Rate Range | ± 5% or ± 2 bpm, whichever is greater |
| Update Interval | Upon measurement completion |
| Measurement Time | 30 seconds (typical) < 135 seconds (maximum) |
| Initial Cuff Pressure | 160 mmHg (user-selectable) |
| Repeated Cuff Pressure | Previous systolic + 40 mmHg |
| Static Cuff Pressure Accuracy | ± 3 mmHg |
| Overpressure Cutoff | 290 ± 3 mmHg (normal means), 300 ± 10 mmHg (back-up) |
| Measurement Modes | Single Measurement or Auto (Interval) Measurement |
| Auto Measurement Settings | Meets ISO 21647 Clause 101.1 (Tables 101 and 105): ± (volume fraction of 0.43% + 8% of gas level) |
| Masimo ISA™ CO₂ / Capnography | |
| Method | Sidestream (Non-dispersive IR) |
| Units | mmHg |
| Parameters | ETCO ₂ , FICO ₂ , RRc |
| CO ₂ Measurement Range | ETCO ₂ & FICO ₂ : 0 to 150 mmHg |
| CO ₂ Measurement Accuracy | As measured with dry single gases: 0 to 15 vol%: ± (0.2 vol% + 2% of reading) 15 to 25 vol%: unspecified |
| CO ₂ Resolution | 1 mmHg |
| RRc (Resp. Rate) Measurement Range | 0 to 150 ± 1 bpm |

| | |
|---|---|
| RRc Resolution | 1 bpm |
| Barometric Pressure Range | 525 to 1200 hPa |
| Barometric Pressure Compensation | Automatic |
| Report Interval | Per breath |
| Flow Rate | 50±10 ml/min |
| Warm-up Time Required to Meet Accuracy Specifications | < 10 seconds (concentrations reported and full accuracy) |
| Total System Response Time | < 3 seconds (with 2m Nomoline sampling line) |
| Drift of Measurement Accuracy | Complies with EN ISO 21647:2004 standard |
| Measurement Accuracy for Gas Mixture | Complies with EN ISO 21647:2004 standard |
| Measurement Accuracy in the Presence of Interfering Gases | Complies with EN ISO 21647:2004 standard |
| Masimo IRMA™ CO₂ / Capnography | |
| Method | Mainstream |
| Units | mmHg |
| Parameters | ETCO ₂ , FICO ₂ , RRc |
| CO ₂ Measurement Range | ETCO ₂ & FICO ₂ : 0 to 150 mmHg |
| CO ₂ Measurement Accuracy | Dry single gases at 22 ± 5°C and 1013 ± 40 hPa 0 to 15 vol%: ± (0.2 vol% + 2% of reading) 15 to 25 vol%: unspecified All conditions ± (0.3 kPa + 4% of reading) |
| CO ₂ Resolution | 1 mmHg |
| RRc (Resp. Rate) Measurement Range | 0 to 150 bpm. RRc is displayed after 3 breaths and the average value is updated every breath. |
| RRc Resolution | 1 bpm |
| Barometric Pressure Range | 525 to 1200 hPa |
| Barometric Pressure Compensation | Automatic |
| Report Interval | Per breath |
| Warm-up Time Required to Meet Accuracy Specifications | < 10 seconds (concentrations reported and full accuracy) |
| Total System Response Time < 10 seconds (concentrations reported and full accuracy) | < 1 second |

| | |
|--|---|
| Drift of Measurement Accuracy | Complies with EN ISO 21647:2004 standard |
| Measurement Accuracy for Gas Complies with EN ISO 21647:2004 standard Mixture | Complies with EN ISO 21647:2004 standard |
| Measurement Accuracy in the Presence of Interfering Gases | Complies with EN ISO 21647:2004 standard |
| Invasive Blood Pressure | |
| Transducer Type | Strain gauge |
| Transducer Excitation Voltage | 5.00 VDC \pm 1 % |
| Frequency Response | 0-12 Hz |
| Measurement Units | mmHg |
| Parameters | Diastolic, Systolic, Mean for all except Mean-only for CVP |
| Measurement Range | -50 to 300 mmHg |
| Measurement Accuracy | \pm 1 mmHg or \pm 1 %, whichever greater |
| IBP Resolution | 1 mmHg |
| Pulse Rate Measurement Range | 30 - 250 bpm |
| Pulse Rate Accuracy | \pm 2 bpm or \pm 2 %, whichever greater |
| Numeric Update Rate | Every 3 seconds |
| Temperature | |
| Compatibility | YSI 400-series probes |
| Measurement Mode | Direct (as defined in ISO 80601 -2-56) |
| Input Connector | 2-pin connector |
| Display Units | $^{\circ}$ F and $^{\circ}$ C (user-selectable) |
| Measurement Resolution | 0.1 $^{\circ}$ F (0.1 $^{\circ}$ C) |
| Measurement Range | 41.0 to 122.0 $^{\circ}$ F (5.0 to 50.0 $^{\circ}$ C) |
| Measurement Accuracy | \pm 0.2 $^{\circ}$ F (\pm 0.1 $^{\circ}$ C) plus probe tolerance |
| Transient Response | Within 30 seconds from 25 to 27 $^{\circ}$ C |
| | |